No.: 13 61 /Purchase Office/AMD/ Ultrasonic Probe /2014

Telephone : 91-532-227+1056

Fax : 0532-2545341, 2545822

GRAM : MNNIT

Email

Website : http://www.mnnit.ac.in

: purchasemnnit@gmail.com,

fip@mnnit.ac.in



मोतीलाल नेहरू राष्ट्रीय प्रौद्योगिकी संस्थान इलाहाबाद,

इलाहाबाद-211004 (भारत)

Motilal Nehru National Institute of Technology Allahabad

Allahabad-211004 (India)

An Institute of National Importance as Declared by NIT Act, GOI, 2

Dated: 18 /09/2014

fip@mnit.ac.in							
	INVITATION FOR SUPPLY OF ULTRASONIC PROBE (For Contracts value estimated to cost less than Rs.25 lakhs)						
	1	Closing Date/Time for submission of Quotations	14/10/2014; 16:30 Hrs.				
	2	Opening Date/Time of Quotations	14/10/2014; 17:00 Hrs.				
	3	Quotations to be submitted at	The Office of Faculty-In-charge (Purchase) M.N.N.I.T. Allahabad, Allahabad-211 004 (U.P.)				
	4	Place of opening of quotations	The Office of Faculty-In-charge (Purchase) M.N.N.I.T. Allahabad, Allahabad-211 004 (U.P.)				
	5	Tender Fee	Rs. 200.00 (Non-refundable), in form of a Demand Draft drawn in favor of DIRECTOR, MNNIT Allahabad.				
	6	Earnest Money Deposit (EMD)	Rs. 14,000/- (Rs. Fourteen Thousand only), in form of FDR/Bank guarantee, from any of the Commercial Banks, drawn in favour of "Director, MNNIT Allahabad" payable at Allahabad, and valid for a period of 45 days beyond the final bid validity period				
1							

To,

Dear Sir

You are invited to submit your most competitive quotation for the following items. The quotation should be sent directly to the undersigned under Sealed Cover marked "Quotation Reference No., Date", and "The Due Date:

SI. No.	Description of item with Specification	Quantity
1.	ULTRASONIC PROBE	01 No.
	(Detailed specifications attached at Annexure-B)	

The sealed tenders must be submitted in the Office of the Faculty In-charge (Purchase), on or before **due date & time as mentioned above** in the office of the undersigned, in the presence of bidder's authorized representatives, who choose to attend and shall sign a register evidencing their attendance.

The tender/quotations NOT submitted in the prescribed format or Incomplete in any respect will be outrightly rejected.

Faculty in-Charge, Purchase

MNNIT Allahabad

GENERAL TERMS & CONDITION (GCC)

1. Bid Price

- (a) The contract shall be for the full quantity as described above. Corrections, if any, shall be made by crossing out, initialing, dating and rewriting.
- (b) All duties, taxes and other levies payable by the contractor under the contract shall be included in the total price.
- (c) The rates quoted by the bidder shall be fixed for the duration of the contract and shall not be subject to adjustment on any account.
- (d) The prices should be quoted in Indian Rupees only for indigenous items. For imported items/Equipments, the prices should be quoted in foreign currency.
- (e) Each bidder shall submit only one quotation.

2. Validity of Quotation

Quotation shall remain valid for a period not less than 90 days after the deadline date specified for submission.

3. Evaluation of Quotations

- (a) The purchaser will evaluate and compare the quotations determined to be substantially responsive, i.e., which
 - (i) are properly signed; and
 - (ii) conform to the terms and conditions, and specifications.
 - (iii) will be with tender fees, and EMD required, if any.
- (b) The Quotations would be evaluated separately for each item.

4. Award of Contract

The Purchaser will award the contract to the bidder whose quotation has been determined to be substantially responsive and who has offered the lowest evaluated quotation price.

- Notwithstanding the above, the Purchaser reserves the right to accept or reject any quotations and to cancel the bidding process and reject all quotations at any time prior to the award of contract.
- The bidder whose bid is accepted will be notified for the award of contract by the Purchaser prior to expiration of the quotation validity period. The terms of the accepted offer shall be incorporated in the purchase order if falls under Govt. purchase rules, i.e., GFR-2005

5. Formats and Signing of Bid

- Each & every pages of the submitted bid must carry the page numbers.
- The bidders are not allowed to make additional and alteration in the tender document, any additions and alteration in the tender document, any additions and alterations shall be at bidder's risk. Conditional/modified tender are liable to be rejected.
- The tender/quotations NOT submitted in the prescribed format or Incomplete in any respect will be outrightly rejected.
- If the bid submitted without compliance sheet details of equipments the bid will be outrightly rejected.
- The bid shall be typed or written in indelible ink and shall be signed by the Bidder/Tenderer or a person or persons duly authorized to bind the Bidder/Tenderer to the Contract. All pages of the bid, except for unamended printed literature, shall be initialed by the person or persons signing the bid.
- Any interlineations, erasure or overwriting shall be valid only if they are initiated by the persons or persons signing the bid.
- The Bidder shall furnish information as described in the Form of Bid on commissions or gratuities, if any, paid or
 to be paid to agents relating to this Bid, and to contract execution if the Bidder is awarded the contract.
- 6. Payment condition: 100% Payment shall be made only after satisfactory delivery and installation of the goods.

 Advance payment is not permitted.
- Normal Commercial warranty/ guarantee shall be applicable to the supplied goods.
- 3. Payment will be cleared through e-payment mode (Format of e-payment form given at Annexure-D)

INSTRUCTIONS/GUIDELINES TO BIDDERS (ITB)

- Bidders who had supplied the Government departments, PSUs/NITs/IITs and reputed Institutions and Organizations should only apply.
- 2. The bidder should enclose the cliental list with contact address along with Phone/Fax for reference, for the quoted item
- 3. Please quote the prices of F.O.R. Motilal Nehru National Institute of Technology Allahabad, Allahabad.
- 4. The Bidders are requested to submit their Quotation with EMD in the form of Demand Draft/ Bank Guarantee/ Short term Deposit of the cost of equipments/items in favor of "The Director, MNNIT, Allahabad" payable at Allahabad. Quotation without EMD will be summarily rejected. EMD's of all Bidders will be returned back after the purchase order /contract agreement is issued to successful bidder.
- Successful Bidders must have to submit the performance security @ 10% of the purchase order or contract value and it will be retained up to the warranty period.
- 6. The bidder should enclose relevant documents wherever necessary to substantiate his eligibility.
- The Institute is eligible for and can provide necessary documentation for Custom/Excise and Octroi exemption on demand, if permitted under rule.
- 8. Motilal Nehru National Institute of Technology Allahabad, Allahabad, reserves the right to reject, accept any or whole or part of any of the tender without assigning any reason thereof and no claim will be heard. In case of dispute, the decision of The Director, MNNIT, and Allahabad will be final and binding.
- 9. If the successful bidder is not able to supply the goods items within the delivery date specified in the purchase order, the purchase order will be automatically treated as cancelled after expiry of delivery date period. The vendor will have to submit an undertaking in this regard that this condition is acceptable to him. In case of extra ordinary circumstances the vendors must send a request for extension of validity of purchase orders, with proper justification prior to the expiry of validity date for consideration.
- 10. In case of sophisticated equipments, an on-site training needs to be provided with no extra cost.
- 11. The quotation must be accompanied by leaflets/operation manuals/books etc.
- 12. Wherever applicable the vendors must quote the branded /desired items.

CHECK LIST (ON THE LETTER HEAD OF THE BIDDER)

The Bid will be the compilation of following documents, along with required supporting documents. No document in support of

minimum eligibility criteria will be accepted / entertained after opening of tender.

SI. No	Documents		
1.	Cover letter by bidder (On the Letter Head of the Bidder)		
2.	Tender Fee		
3.	Earnest Money Deposit (EMD), if required		
4.	Cliental List for quoted item		
5.	Annexure-A: Bid Proposal Sheet (On the Letter Head of the Bidder)		
6.	Annexure-B: Technical Specifications of Equipment (On the Letter Head of the Bidder)		
7.	Annexure-C: Compliance of bidder with reference to the equipment (On the Letter Head of the Bidder)		
8.	Annexure-D : PRICE BID (On the Letter Head of the Bidder)		
9.	Annexure-E: PROFORMA FOR DIRECT PAYMENT/TRANSFER TO BANK ACCOUNT BY MNNIT ALLAHABAD		

ANNEXURE-A

BID PROPOSAL SHEET (ON THE LETTER HEAD OF THE BIDDER)

To

Faculty In-charge (Purchase) MNNIT Allahabad, Allahabad

Subject: Supply of "ULTRASONIC PROBE" at MNNIT Allahabad, Allahabad.

Dear Sir,

We, the undersigned Tenderers, having read and examined in detail the specifications as specified in this document in respect of Supply of **ULTRASONIC PROBE** at MNNIT Allahabad, Allahabad do hereby propose to supply the required products and services.

Tender No				
Tender Fees submitted		YES / NO	YES / NO (Please strike off whatever is	
Amount	Mode	Date of Issue	Name of Bank	Valid up to
	Demand Draft			
EMD submitted		YES / NO	(Please strike off whate	ver is not applicable)
Amount	Mode	Date of issue	Name of Bank	Valid up to
· · · · · ·	Demand Draft			

- (i) ADDITIONAL PURCHASEWORK ORDER: We understand that the MNNIT Allahabad, Allahabad, in case of the requirements may also place repeat purchase order/work order on the company. In such cases, we shall accept and execute all the purchase/work order placed on us by MNNIT Allahabad, Allahabad.
- (ii) **BID PRICING:** We further declare that the prices stated in our proposal are in accordance with your Terms & Conditions in the bidding document. We further understand that the quantities as specified in this Tender may increase or decrease at the time of Award of Purchase Order as per the requirements of MNNIT Allahabad, Allahabad.
- (iii) QUALIFYING DATA: We confirm that we satisfy the qualifying criteria and have attached the requisite documents as documentary proofs. In case you require any further information/documentary proof in this regard during evaluation of our bid, we agree to furnish the same in time to your satisfaction.
- (iv) CONTRACT PERFORMANCE SECURITY: We hereby declare that in case the contract is awarded to us, we shall submit the performance Guarantee Bond in the form of Bank Guarantee for the amount of 10% of the total order value.
- (v) PAYMENT TERMS: We hereby declare that in case the contract is awarded to us, we agree with payment terms specified in the tender documents.
- (vi) CERTIFICATE AND DECLARATION:
 - a) I/We certify that no addition/modification/alteration has been made in the Original Tender Document. If at any stage addition /modification /alteration is noticed in the Original Document, I/We will abide by the terms and conditions contained in the original tender document, failing which MNNIT Allahabad reserves the right to reject the tender and/or cancel the contract
 - b) It has been certified that all information provided in tender form is true and correct to the best of my knowledge and belief. We hereby declare that our proposal is made in good faith, without collusion or fraud. No forged /tampered document(s) are produced with tender form for gaining unlawful advantage. We understand that MNNIT Allahabad is authorized to make enquiry to establish the facts claimed and obtained confidential reports from clients.
 - c) In case it is established that any information provided by us is false / misleading or in the circumstances where it is found that we have made any wrong claims. Further MNNIT Allahabad is also authorized to blacklist our firm/company/agency and debar us in participating in any tender/bid in future.
 - d) I / We assure the Institute that neither I /We, nor any of my /our workers, will do any act which is improper / illegal during the execution in case the tender is awarded to us.
 - e) I / We assure the Institute that I / We will NOT be outsourcing any work specified in the tender document, to any other firm.
 - f) Neither I / We, nor anybody on my /our behalf will indulge in any corrupt activities /practices in my /our dealing with the Institute.
 - g) Our Firm / Company / Agency is not been blacklisted or banned by any Govt. Department, PSU, University, Autonomous Institute or any other Govt. Organization.
 - h) I/We certify that, I have understood all the terms & conditions (GCC and ITB etc.), as indicated in enquiry of the tender document, and hereby accept all the same completely.
 - i) I/We, further certify that I/We, possess all the statutory /non-statutory registrations, permissions, approvals, etc., from the Competent Authority for providing the requisite services,
 - j) We understand that you are not bound to accept the lowest or any bid you may receive.
 - k) I/We hereby declare that this tender on acceptance communicated by you shall constitute a valid and binding contract between us.
 - I) I/We certify that the submitted quotation is duly paginated and contains from page no. 1 to

Technical Specifications of 'ULTRASONIC PROBE'
(ON THE LETTER HEAD OF THE BIDDER)

SI. No.	Description		
1.	Ultrasonicator with 500 W or more for High Intensity Ultra Sonic Liquid		
2.	Processor, 250 ul to liters, and 20KHz		
3. Ultrasonicator should have Automatic tuning, energy (joules), over load protection circuit			
It should monitor - digital wattmeter, and have Integrated Temperature Controller			
5.	Temperature Probe to measure samples temperature up to 100deg C		
6.	Equipment should have micro- processor - based, programmable, LCD screen with user prompts, Converter should be sealed and display real time		
7.	Equipment should have 10 or more programs storage capability.		
8.	It can control independently processing time from 1 second to 10 hours		
9.	Ultrasonicator should have Automatic Amplitude Compensation to ensure uniform probe amplitude		
10.			
11.	Equipment should be supplied with essential accessories, Convertible cable length should be 5' or		
	more		
12.	Dimension should be approximately 230x190x330mm		
13.	Assorted types probes and tips for different volumes as optional may be quoted as optional		
14.	Ultrasonic Bath of dimensions: 240X140X100mm with 3.0 Ltr capacity (Stainless Steel tank and		
	housing, timer 1 – 15 min and infinity, Transducers: 1 PZT – large area, Generator should be Overloa		
	protected, constant power, Full wave, Frequency: 35kHz automatic frequency control, Mains connection: 220-240V,50/60Hz, Thermostatic Control 30 - 80°C, Furnace has built -in heating facility		
15.	Ultarsonicators should have sound abetting system to prevent ultrasonic waves		
16.	Accessories like adjustable platform, chamber etc should be provided.		
17.	The vendor should have good service support.		
18.	The vendor should provide free demonstration, application support and training facility.		
19.	Also provide list of users in India.		
20.	Warranty: 3 year		

ANNEXURE-C

Technical Compliance of 'Ultrasonic Probe' (ON THE LETTER HEAD OF THE BIDDER)

No. the requirement (Yes/ No) 1. Ultrasonicator with 500 W or more for High Intensity Ultra Sonic Liquid 2. Processor, 250 ul to liters, and 20KHz 3. Ultrasonicator should have Automatic tuning, energy (joules), over load protection circuit 4. It should monitor - digital wattmeter, and have Integrated Temperature Controller 5. Temperature Probe to measure samples temperature up to 100deg C 6. Equipment should have micro- processor - based, programmable, LCD screen with user prompts, Converter should be sealed and display real time 7. Equipment should have 10 or more programs storage capability. 8. It can control independently processing time from 1 second to 10 hours 9. Ultrasonicator should have Automatic Amplitude Compensation to ensure uniform probe amplitude 10. Standard probe ½" (13mm) and 19mm Solid Probe or any other size with replaceable tip 11. Equipment should be supplied with essential accessories, Convertible cable length should be 5' or more 12. Dimension should be approximately 230x190x330mm 13. Assorted types probes and tips for different volumes as optional may be quoted as optional 14. Ultrasonic Bath of dimensions: 240X140X100mm with 3.0 Ltr capacity (Stainless Steel tank and housing, timer 1 – 15 min and infinity, Transducers: 1 PZT – large area, Generator should be Overload protected, constant power, Full wave, Frequency: 35kHz automatic frequency control, Mains connection: 220-240V,50/60Hz, Thermostatic Control 30 - 80°C, Furnace has built -in heating facility 15. Ultrasonicators should have sound abetting system to prevent ultrasonic waves 16. Accessories like adjustable platform, chamber etc should be provided. 17. The vendor should provide free demonstration, application support and training facility. 18. House of the series in India. 19. Warranty: 3 year	SI.	SI. Description Whether meet Quote				
 Processor, 250 ul to liters, and 20KHz Ultrasonicator should have Automatic tuning, energy (joules), over load protection circuit It should monitor - digital wattmeter, and have Integrated Temperature Controller Temperature Probe to measure samples temperature up to 100deg C Equipment should have micro- processor - based, programmable, LCD screen with user prompts, Converter should be sealed and display real time Equipment should have 10 or more programs storage capability. It can control independently processing time from 1 second to 10 hours Ultrasonicator should have Automatic Amplitude Compensation to ensure uniform probe amplitude Standard probe ½" (13mm) and 19mm Solid Probe or any other size with replaceable tip Equipment should be supplied with essential accessories, Convertible cable length should be 5' or more Dimension should be approximately 230x190x330mm Assorted types probes and tips for different volumes as optional may be quoted as optional Ultrasonic Bath of dimensions: 240X140X100mm with 3.0 Ltr capacity (Stainless Steel tank and housing, timer 1 – 15 min and infinity, Transducers: 1 PZT – large area, Generator should be Overload protected, constant power, Full wave, Frequency: 35kHz automatic frequency control, Mains connection: 220-240V,50/60Hz, Thermostatic Control 30 - 80°C, Furnace has built -in heating facility Ultrasonicators should have sound abetting system to prevent ultrasonic waves Accessories like adjustable platform, chamber etc should be provided. The vendor should have good service support. The vendor should have sin India. 				Specifications		
 Ultrasonicator should have Automatic tuning, energy (joules), over load protection circuit It should monitor - digital wattmeter, and have Integrated Temperature Controller Temperature Probe to measure samples temperature up to 100deg C Equipment should have micro- processor - based, programmable, LCD screen with user prompts, Converter should be sealed and display real time Equipment should have 10 or more programs storage capability. It can control independently processing time from 1 second to 10 hours Ultrasonicator should have Automatic Amplitude Compensation to ensure uniform probe amplitude Standard probe ½" (13mm) and 19mm Solid Probe or any other size with replaceable tip Equipment should be supplied with essential accessories, Convertible cable length should be 5' or more Dimension should be approximately 230x190x330mm Assorted types probes and tips for different volumes as optional may be quoted as optional Ultrasonic Bath of dimensions: 240X140X100mm with 3.0 Ltr capacity (Stainless Steel tank and housing, timer 1 – 15 min and infinity, Transducers: 1 PZT – large area, Generator should be Overload protected, constant power, Full wave, Frequency: 35kHz automatic frequency control, Mains connection: 220-240V,50/60Hz, Thermostatic Control 30 - 80°C, Furnace has built -in heating facility Ultrasonicators should have sound abetting system to prevent ultrasonic waves Accessories like adjustable platform, chamber etc should be provided. The vendor should have good service support. The vendor should have good service support. Also provide list of users in India. 	1.	Ultrasonicator with 500 W or more for High Intensity Ultra Sonic Liquid				
protection circuit It should monitor - digital wattmeter, and have Integrated Temperature Controller Temperature Probe to measure samples temperature up to 100deg C Equipment should have micro- processor - based, programmable, LCD screen with user prompts, Converter should be sealed and display real time Equipment should have 10 or more programs storage capability. It can control independently processing time from 1 second to 10 hours Ultrasonicator should have Automatic Amplitude Compensation to ensure uniform probe amplitude 10. Standard probe ½" (13mm) and 19mm Solid Probe or any other size with replaceable tip 11. Equipment should be supplied with essential accessories, Convertible cable length should be 5' or more 12. Dimension should be approximately 230x190x330mm 13. Assorted types probes and tips for different volumes as optional may be quoted as optional 14. Ultrasonic Bath of dimensions: 240X140X100mm with 3.0 Ltr capacity (Stainless Steel tank and housing, timer 1 – 15 min and infinity, Transducers: 1 PZT – large area, Generator should be Overload protected, constant power, Full wave, Frequency: 35kHz automatic frequency control, Mains connection: 220-240V,50/60Hz, Thermostatic Control 30 - 80°C, Furnace has built -in heating facility 15. Ultrasonicators should have sound abetting system to prevent ultrasonic waves 16. Accessories like adjustable platform, chamber etc should be provided. 17. The vendor should have good service support. 18. The vendor should have good service support. 19. Also provide list of users in India.	2.	Processor, 250 ul to liters, and 20KHz				
Controller Temperature Probe to measure samples temperature up to 100deg C Equipment should have micro- processor - based, programmable, LCD screen with user prompts, Converter should be sealed and display real time Equipment should have 10 or more programs storage capability. It can control independently processing time from 1 second to 10 hours Ultrasonicator should have Automatic Amplitude Compensation to ensure uniform probe amplitude Standard probe ½" (13mm) and 19mm Solid Probe or any other size with replaceable tip Equipment should be supplied with essential accessories, Convertible cable length should be 5' or more Dimension should be approximately 230x190x330mm Assorted types probes and tips for different volumes as optional may be quoted as optional Ultrasonic Bath of dimensions: 240X140X100mm with 3.0 Ltr capacity (Stainless Steel tank and housing, timer 1 – 15 min and infinity, Transducers: 1 PZT – large area, Generator should be Overload protected, constant power, Full wave, Frequency: 35kHz automatic frequency control, Mains connection: 220-240V,50/60Hz, Thermostatic Control 30 - 80°C, Furnace has built -In heating facility Ultarsonicators should have sound abetting system to prevent ultrasonic waves 16. Accessories like adjustable platform, chamber etc should be provided. 17. The vendor should provide free demonstration, application support and training facility. 19. Also provide list of users in India.	3.					
6. Equipment should have micro- processor - based, programmable, LCD screen with user prompts, Converter should be sealed and display real time 7. Equipment should have 10 or more programs storage capability. 8. It can control independently processing time from 1 second to 10 hours 9. Ultrasonicator should have Automatic Amplitude Compensation to ensure uniform probe amplitude 10. Standard probe ½" (13mm) and 19mm Solid Probe or any other size with replaceable tip 11. Equipment should be supplied with essential accessories, Convertible cable length should be 5' or more 12. Dimension should be approximately 230x190x330mm 13. Assorted types probes and tips for different volumes as optional may be quoted as optional 14. Ultrasonic Bath of dimensions: 240X140X100mm with 3.0 Ltr capacity (Stainless Steel tank and housing, timer 1 – 15 min and infinity, Transducers: 1 PZT – large area, Generator should be Overload protected, constant power, Full wave, Frequency: 35kHz automatic frequency control, Mains connection: 220-240V,50/60Hz, Thermostatic Control 30 - 80°C, Furnace has built -in heating facility 15. Ultrasonicators should have sound abetting system to prevent ultrasonic waves 16. Accessories like adjustable platform, chamber etc should be provided. 17. The vendor should have good service support. 18. Also provide list of users in India.	4.					
screen with user prompts, Converter should be sealed and display real time 7. Equipment should have 10 or more programs storage capability. 8. It can control independently processing time from 1 second to 10 hours 9. Ultrasonicator should have Automatic Amplitude Compensation to ensure uniform probe amplitude 10. Standard probe ½" (13mm) and 19mm Solid Probe or any other size with replaceable tip 11. Equipment should be supplied with essential accessories, Convertible cable length should be 5' or more 12. Dimension should be approximately 230x190x330mm 13. Assorted types probes and tips for different volumes as optional may be quoted as optional 14. Ultrasonic Bath of dimensions: 240X140X100mm with 3.0 Ltr capacity (Stainless Steel tank and housing, timer 1 – 15 min and infinity, Transducers: 1 PZT – large area, Generator should be Overload protected, constant power, Full wave, Frequency: 35kHz automatic frequency control, Mains connection: 220-240V,50/60Hz, Thermostatic Control 30 - 80°C, Furnace has built -in heating facility 15. Ultrasonicators should have sound abetting system to prevent ultrasonic waves 16. Accessories like adjustable platform, chamber etc should be provided. 17. The vendor should have good service support. 18. The vendor should provide free demonstration, application support and training facility. 19. Also provide list of users in India.	5.	Temperature Probe to measure samples temperature up to 100deg C				
 Equipment should have 10 or more programs storage capability. It can control independently processing time from 1 second to 10 hours Ultrasonicator should have Automatic Amplitude Compensation to ensure uniform probe amplitude Standard probe ½" (13mm) and 19mm Solid Probe or any other size with replaceable tip Equipment should be supplied with essential accessories, Convertible cable length should be 5' or more Dimension should be approximately 230x190x330mm Assorted types probes and tips for different volumes as optional may be quoted as optional Ultrasonic Bath of dimensions: 240X140X100mm with 3.0 Ltr capacity (Stainless Steel tank and housing, timer 1 – 15 min and infinity, Transducers: 1 PZT – large area, Generator should be Overload protected, constant power, Full wave, Frequency: 35kHz automatic frequency control, Mains connection: 220-240V,50/60Hz, Thermostatic Control 30 - 80°C, Furnace has built -in heating facility Ultarsonicators should have sound abetting system to prevent ultrasonic waves Accessories like adjustable platform, chamber etc should be provided. The vendor should have good service support. The vendor should provide free demonstration, application support and training facility. Also provide list of users in India. 	6.					
 8. It can control independently processing time from 1 second to 10 hours 9. Ultrasonicator should have Automatic Amplitude Compensation to ensure uniform probe amplitude 10. Standard probe ½" (13mm) and 19mm Solid Probe or any other size with replaceable tip 11. Equipment should be supplied with essential accessories, Convertible cable length should be 5' or more 12. Dimension should be approximately 230x190x330mm 13. Assorted types probes and tips for different volumes as optional may be quoted as optional 14. Ultrasonic Bath of dimensions: 240X140X100mm with 3.0 Ltr capacity (Stainless Steel tank and housing, timer 1 – 15 min and infinity, Transducers: 1 PZT – large area, Generator should be Overload protected, constant power, Full wave, Frequency: 35kHz automatic frequency control, Mains connection: 220-240V,50/60Hz, Thermostatic Control 30 - 80°C, Furnace has built -in heating facility 15. Ultarsonicators should have sound abetting system to prevent ultrasonic waves 16. Accessories like adjustable platform, chamber etc should be provided. 17. The vendor should have good service support. 18. The vendor should provide free demonstration, application support and training facility. 19. Also provide list of users in India. 	7.					
 9. Ultrasonicator should have Automatic Amplitude Compensation to ensure uniform probe amplitude 10. Standard probe ½" (13mm) and 19mm Solid Probe or any other size with replaceable tip 11. Equipment should be supplied with essential accessories, Convertible cable length should be 5' or more 12. Dimension should be approximately 230x190x330mm 13. Assorted types probes and tips for different volumes as optional may be quoted as optional 14. Ultrasonic Bath of dimensions: 240X140X100mm with 3.0 Ltr capacity (Stainless Steel tank and housing, timer 1 – 15 min and infinity, Transducers: 1 PZT – large area, Generator should be Overload protected, constant power, Full wave, Frequency: 35kHz automatic frequency control, Mains connection: 220-240V,50/60Hz, Thermostatic Control 30 - 80°C, Furnace has built -in heating facility 15. Ultarsonicators should have sound abetting system to prevent ultrasonic waves 16. Accessories like adjustable platform, chamber etc should be provided. 17. The vendor should have good service support. 18. The vendor should provide free demonstration, application support and training facility. 19. Also provide list of users in India. 						
 Standard probe ½" (13mm) and 19mm Solid Probe or any other size with replaceable tip Equipment should be supplied with essential accessories, Convertible cable length should be 5' or more Dimension should be approximately 230x190x330mm Assorted types probes and tips for different volumes as optional may be quoted as optional Ultrasonic Bath of dimensions: 240X140X100mm with 3.0 Ltr capacity (Stainless Steel tank and housing, timer 1 – 15 min and infinity, Transducers: 1 PZT – large area, Generator should be Overload protected, constant power, Full wave, Frequency: 35kHz automatic frequency control, Mains connection: 220-240V,50/60Hz, Thermostatic Control 30 - 80°C, Furnace has built -in heating facility Ultrasonicators should have sound abetting system to prevent ultrasonic waves Accessories like adjustable platform, chamber etc should be provided. The vendor should have good service support. The vendor should provide free demonstration, application support and training facility. Also provide list of users in India. 		Ultrasonicator should have Automatic Amplitude Compensation to ensure				
 Equipment should be supplied with essential accessories, Convertible cable length should be 5' or more Dimension should be approximately 230x190x330mm Assorted types probes and tips for different volumes as optional may be quoted as optional Ultrasonic Bath of dimensions: 240X140X100mm with 3.0 Ltr capacity (Stainless Steel tank and housing, timer 1 – 15 min and infinity, Transducers: 1 PZT – large area, Generator should be Overload protected, constant power, Full wave, Frequency: 35kHz automatic frequency control, Mains connection: 220-240V,50/60Hz, Thermostatic Control 30 - 80°C, Furnace has built -in heating facility Ultarsonicators should have sound abetting system to prevent ultrasonic waves Accessories like adjustable platform, chamber etc should be provided. The vendor should have good service support. The vendor should provide free demonstration, application support and training facility. Also provide list of users in India. 	10.	Standard probe 1/2" (13mm) and 19mm Solid Probe or any other size with				
 12. Dimension should be approximately 230x190x330mm 13. Assorted types probes and tips for different volumes as optional may be quoted as optional 14. Ultrasonic Bath of dimensions: 240X140X100mm with 3.0 Ltr capacity (Stainless Steel tank and housing, timer 1 – 15 min and infinity, Transducers: 1 PZT – large area, Generator should be Overload protected, constant power, Full wave, Frequency: 35kHz automatic frequency control, Mains connection: 220-240V,50/60Hz, Thermostatic Control 30 - 80°C, Furnace has built -in heating facility 15. Ultarsonicators should have sound abetting system to prevent ultrasonic waves 16. Accessories like adjustable platform, chamber etc should be provided. 17. The vendor should have good service support. 18. The vendor should provide free demonstration, application support and training facility. 19. Also provide list of users in India. 	11.	Equipment should be supplied with essential accessories, Convertible cable				
 Assorted types probes and tips for different volumes as optional may be quoted as optional Ultrasonic Bath of dimensions: 240X140X100mm with 3.0 Ltr capacity (Stainless Steel tank and housing, timer 1 – 15 min and infinity, Transducers: 1 PZT – large area, Generator should be Overload protected, constant power, Full wave, Frequency: 35kHz automatic frequency control, Mains connection: 220-240V,50/60Hz, Thermostatic Control 30 - 80°C, Furnace has built -in heating facility Ultrasonicators should have sound abetting system to prevent ultrasonic waves Accessories like adjustable platform, chamber etc should be provided. The vendor should have good service support. The vendor should provide free demonstration, application support and training facility. Also provide list of users in India. 	12.		1			
 14. Ultrasonic Bath of dimensions: 240X140X100mm with 3.0 Ltr capacity (Stainless Steel tank and housing, timer 1 – 15 min and infinity, Transducers: 1 PZT – large area, Generator should be Overload protected, constant power, Full wave, Frequency: 35kHz automatic frequency control, Mains connection: 220-240V,50/60Hz, Thermostatic Control 30 - 80°C, Furnace has built -in heating facility 15. Ultarsonicators should have sound abetting system to prevent ultrasonic waves 16. Accessories like adjustable platform, chamber etc should be provided. 17. The vendor should have good service support. 18. The vendor should provide free demonstration, application support and training facility. 19. Also provide list of users in India. 		Assorted types probes and tips for different volumes as optional may be				
waves 16. Accessories like adjustable platform, chamber etc should be provided. 17. The vendor should have good service support. 18. The vendor should provide free demonstration, application support and training facility. 19. Also provide list of users in India.	14.	Ultrasonic Bath of dimensions: 240X140X100mm with 3.0 Ltr capacity (Stainless Steel tank and housing, timer 1 – 15 min and infinity, Transducers: 1 PZT – large area, Generator should be Overload protected, constant power, Full wave, Frequency: 35kHz automatic frequency control, Mains connection: 220-240V,50/60Hz, Thermostatic Control 30 - 80°C, Furnace has built -in				
 17. The vendor should have good service support. 18. The vendor should provide free demonstration, application support and training facility. 19. Also provide list of users in India. 	15.					
 17. The vendor should have good service support. 18. The vendor should provide free demonstration, application support and training facility. 19. Also provide list of users in India. 	16.	Accessories like adjustable platform, chamber etc should be provided.				
 18. The vendor should provide free demonstration, application support and training facility. 19. Also provide list of users in India. 						
19. Also provide list of users in India.		The vendor should provide free demonstration, application support and				
	19.					
			>			

PRICE BID (ON THE LETTER HEAD OF THE BIDDER)

SI. No.	Item Description	Qty	Offered Make/ Model	U/ Price	T/Price In Figures	(in Rs.) In Words
1.	ULTRASONIC PROBE	01 No.				
	(Detailed specifications attached at Annexure-B)					

We agree to supply the above goods in accordance with the technical specifications for a total contract price of within the period specified in the Invitation for Quotations.

Signature and Seal of the Manufacturer/Bidder

Note:

1. Bidders should provide copies of original Memorandum and Articles of Association, defining the constitution of legal status, place of registration and place of business of the company.

ANNEXURE-E PROFORMA FOR DIRECT PAYMENT/TRANSFER TO BANK ACCOUNT BY MNNIT ALLAHABAD

SI. No.	Particulars	Information
1.	Firm [Beneficiary] Name	
2.	Code No. allotted by the MNNIT Allahabad [If code is not allotted yet, please enclose a cancelled cheque and copy of PAN card. Cancelled cheque & PAN card is to be submitted only once]	
3.	Complete Bank Account No. of the Firm [beneficiary]. [in case of change in bank account vendor should write to Account Office]	
4.	Bank Name	
5.	Branch Address	
6.	IFSC Code No.	
7.	Permanent Account Number	
8.	Mobile No. [for SMS]	
9.	Email ID [for information]	

We undertake that all information provided above is correct and MNNIT Allahabad will not be responsible in case of any error on the part of firm.

[Seal and Signature of the firm]